

Name: Xiamen Institute of Measurement and Testing

Address: No.265, Andou East Road, Huli District, Xiamen, Fujian, China

Registration No. CNAS L3031

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2024-08-01 Expiry Date: 2030-04-11

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1.Mechanics							
1	*Digital Indicating Weighing Instrument	mass	V.R.of Digital Indicating Weighing Instruments JJG539	4g~60kg	U= (0.03~2) g		
				60kg~1t	U=2g~0.06kg		
				(1~10)t	U=(0.06~1.0)kg		
				(10~100)t	U= (1.0~6) kg		
2	*Non-self Indicating Weight	mass	V.R.of Non-self indication Weighing Instruments JJG14	(1~10)kg	U=(0.4~1.5)g		
				(10~100)kg	U=(1.5~15)g		
				100kg~1t	U=15g~0.15kg		
3	*Analogue Indicating Weighing Instrument	mass	V.R.of Analogue Indicating Weighing Instruments JJG13	(0.25~8)kg	U=(1~4)g		
				(8~50)kg	U=(4~10)g		



No. CNAS L3031

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
4	*Rockwell hardness Testers	Hardness	V.R.of Metallic Rockwell Hardness Testers(Scales A,B,C,D,E,F,G,H,K,N,T) JJG112	(20~70)HRC	$U=0.5\text{HRC}$		
				(85~100)HRBW	$U=0.7\text{HRBW}$		
				(80~88)HRA	$U=0.5\text{HRA}$		
5	*Brinell Hardness Testers	Hardness	V.R.of Metallic Brinell Hardness Testers JJG150	(75~125)HBW	$U_{\text{rel}}=2.5\%$		
				(125~600)HBW	$U_{\text{rel}}=1.9\%$		
6	*Vickers Hardness Testers	Hardness	V.R.of Metallic Vickers Hardness Testing Machines JJG151	(170~225) HV5	$U_{\text{rel}}=5.2\%$		
				(>225~1000) HV5	$U_{\text{rel}}=3.3\%$		
				(>225~1000) HV10	$U_{\text{rel}}=3.3\%$		
				(>225~1000) HV30	$U_{\text{rel}}=3.3\%$		
7	pencil hardness tester	pencil-tip load	Calibration Specification for Pencil Hardness Testers JJF (SH) 007	(500~1000) g	$U=1.6\text{g}$		
		angle		$44^{\circ} \sim 46^{\circ}$	$U=0.24^{\circ}$		
8	*Universal Tension and Compression Testing Machine	Force	V.R.of Tension, Compression and Universal Testing Machines JJG139	0.1N~5MN	$U_{\text{rel}}=0.44\%$		
9	Working Dynamo-meters	Force Indication	V.R.of Working Dynamometers JJG455	(0.01~500)N	$U_{\text{rel}}=0.22\%$		
				500N~300kN	$U_{\text{rel}}=0.44\%$		
10	Hydraulic Jack	Force Indication	V.R.of Hydraulic Jacks JJG621	(10~5000) kN	$U_{\text{rel}}=0.53\%$		
11	Torque Wrenches	Torque	V.R.of Torque Wrenches JJG707	(2~50) Nm	$U_{\text{rel}}=1.4\%$		
				(50~3000) Nm	$U_{\text{rel}}=1.1\%$		



No. CNAS L3031

第 2 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
12	*Calibration Instrument for Torque Wrenches	Torque	V.R.of Calibration Instrument for Torque Wrenches JJG797	(0.1~1000) Nm	<i>U</i> _{rel} =0.15%		
13	Techometer	Rate	V.R.of Techometers JJG105	(20~1000) r/min	<i>U</i> _{rel} =0.3%		
				(1000~40000) r/min	<i>U</i> _{rel} =0.12%		
14	Locomotive Pipette	capacity	V.R.of Locomotive Pipette JJG646	(1~10) μ L	<i>U</i> _{rel} =0.6%		
				(10~10000) μ L	<i>U</i> _{rel} =0.1%		
15	Electromagnetic Flowmeters	Flow	V.R. of Electromagnetic Flowmeters JJG1033	medium: water DN: (15~200) mm (0.1~700) m³/h	<i>U</i> _{rel} =0.06%		
16	Ultrasonic Flowmeters	Flow	V.R. of Ultrasonic Flowmeters JJG 1030	medium: water DN: (15~200) mm (0.1~700) m³/h	<i>U</i> _{rel} =0.07%		
17	Vortex-shedding Flowmeter	Flow	V.R. of Vortex-shedding Flowmeter JJG 1029	(medium: water DN: (15~200) mm (0.1~700) m³/h	<i>U</i> _{rel} =0.1%		
18	Turbine Flowmeter	Flow	V.R.of Turbine Flowmeter JJG1037	medium: water DN: (15~200) mm (0.1~700) m³/h	<i>U</i> _{rel} =0.07%		
2.Chemistry							
1	*Spectrophotometer (UV-VIS)	Wavelength	V. R. of Ultraviolet 、 Visible、Near-Infrared Spectrophotometers JJG178 JJG178	(190~340)nm	<i>U</i> =0.5nm		
				(340~900)nm	<i>U</i> =0.6nm		
		Photometric		A:0% τ ~ 100% τ	<i>U</i> =0.4%		
				B:0% τ ~ 100% τ	<i>U</i> =0.6%		



No. CNAS L3031

第 3 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
2	Laboratory pH Meters	pH	V. R. of Laboratory pH Meters JJG119	Meter:0~14	<i>U</i> =0.01		
				Instrument:3~10	<i>U</i> =0.02		
3	Polarimeter and Polarimetric Saccharimeters	Polarimeter	V. R. of Polarimeter and	-45° ~ +45°	<i>U</i> =0.005°		
		Polarimetric Sacchari	Polarimetric Saccharimeters JJG536	(-20~+105)° Z	<i>U</i> =0.01° Z		
4	Electrolytic Conductivity Meters	Conductivity	V. R. of Electrolytic Conductivity Meters JJG376	Electronic unit:(0.1~1× 10 ⁵) μ S/cm	<i>U</i> _{rel} =0.3%		
				Instrument:(2~2000 μ S/cm	<i>U</i> _{rel} =0.4%		
5	Ionmeter	mV	V. R. of Ionometer JJG757	(0~2000)mV	<i>U</i> =0.7mV		
		pX		Meter:0~14	<i>U</i> =0.01		
				Instrument:3~8	<i>U</i> =0.02		
6	*Gas Chromatograph	Detection limit	V. R. of Gas Chromatograph JJG700	FID:≤0.5ng/s	<i>U</i> _{rel} =4%		
				ECD:≤5pg/mL	<i>U</i> _{rel} =5%		
				FPD:S≤0.5ng/s	<i>U</i> _{rel} =5%		
				FPD:P≤0.1ng/s	<i>U</i> _{rel} =5%		
				NPD:N≤5pg/s	<i>U</i> _{rel} =5%		
				NPD:P≤10pg/s	<i>U</i> _{rel} =5%		
		Sensitivity		TCD:≥ 800mV • ml/mg	<i>U</i> _{rel} =5%		
7	*Liquid chromatograph	Minimum detected concentration	V. R. of Liquid chromatographs JJG705	UVD:≤5× 10 ⁻⁸ g/mL	<i>U</i> _{rel} =8%		
				DAD:≤5× 10 ⁻⁸ g/mL	<i>U</i> _{rel} =8%		



No. CNAS L3031

第 4 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				FAD: $\leq 5 \times 10^{-9}$ g/mL	$U_{\text{rel}}=8\%$		
				RID: $\leq 5 \times 10^{-6}$ g/mL	$U_{\text{rel}}=9\%$		
				ELSD: $\leq 5 \times 10^{-6}$ g/mL	$U_{\text{rel}}=9\%$		
8	*Atomic Absorption Spectrophotometer	Limited calculation	V. R. of Atomic Absorption Spectrophotometer JJG694	flame: $\leq 0.02 \mu\text{g/mL}$	$U=0.005 \mu\text{g/mL}$		
				graphite furnace: $\leq 4\text{pg}$	$U=0.4\text{pg}$		
9	*Ion Chromatograph	Minimum detecting limited concentration	V. R. of Ion Chromatographs JJG823	ECD: $\leq 0.02\text{mg/L}$	$U_{\text{rel}}=13\%$		
				UVD: $\leq 0.02\text{mg/L}$	$U_{\text{rel}}=13\%$		
				ED: $\leq 0.02\text{mg/L}$	$U_{\text{rel}}=13\%$		
10	*Mercury Analyzers	Detection limit	V. R. of Mercury Analyzers JJG548	$\leq 0.1\text{ng}$	$U_{\text{rel}}=10\%$		
11	Carbon Dioxide Infrared Gas Analyzers	Concentration	V. R. of Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzers JJG635	CO: (1~2000) $\mu\text{mol/mol}$	$U_{\text{rel}}=3\%$		
				CO ₂ : (1~25000) $\mu\text{mol/mol}$	$U_{\text{rel}}=2\%$		
12	Formaldehyde Gas Analyzer	Concentration	V. R. of Formaldehyde Gas Analyzers JJG1022	(0.02~1.5) $\mu\text{mol/mol}$	$U_{\text{rel}}=4\%$		
13	*Liquid Chromatography-Mass Spectrometer	S/N ratio	C. S. for Liquid Chromatography-Mass Spectrometers JJF1317	$\geq 10:1$	$U_{\text{rel}}=9\%$		
14	Dissolved Oxygen Meter	Concentration	V. R. of Dissolved Oxygen Meters JJG291	(1~20)mg /L	$U=0.05\text{mg /L}$		
15	*Gas Chromatography-Mass Spectrometers	S/N ratio	C. S. for Gas Chromatography-Mass Spectrometries JJF1164	$\geq 10:1$	$U_{\text{rel}}=15\%$		
16	*Alarmer Detector	Concentration	V. R. of Alarmer Detector of	CH ₄ : (1~100)%LEL	$U_{\text{rel}}=3\%$		



No. CNAS L3031

第 5 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	of Combustible Gas		Combustible Gas JJG693	C ₃ H ₈ :(1~100)%LEL	U _{rel} =3%		
				C ₄ H ₁₀ :(1~100)%LEL	U _{rel} =3%		
17	*Carbon Monoxide Detector	Concentration	V. R. of Carbon Monoxide Detectors JJG915	(1~2000) × 10 ⁻⁶ mol/mol	U _{rel} =4%		
18	*Sulfur Hydrogen Gas Detectors	Concentration	V. R. of Sulfur Hydrogen Gas Detectors JJG695	(1~200) × 10 ⁻⁶ mol/mol	U _{rel} =4%		
19	*Oxygen Analyzer and Oxygen Detector	Concentration	V. R. of Electrochemical Oxygen Meter JJG365	(1~100)%	U _{rel} =2%		
20	*Fourier Transform Infrared Spectrometers	Wavelength	C. S. for Fourier Transform Infrared Spectrometers JJF1319	(4000~400)cm ⁻¹	U=0.3cm ⁻¹		
21	*Inductively Coupled Plasma Mass Spectrometers	Detection limit	C. S. for Quadrupole Inductively Coupled Plasma Mass Spectrometers JJF1159	Be: ≤ 30ng.L ⁻¹ /sup>	U _{rel} =11%		
				In: ≤ 10ng.L ⁻¹	U _{rel} =12%		
				Bi: ≤ 10ng.L ⁻¹	U _{rel} =11%		
22	*Kjeldahl nitrogen determinator	Concentration	C. S. for Elemental Analyzers JJF1321	1%~50%	U=1.2%		
23	*Carbon- sulfur Analyzer	Concentration	V. R. of Carbon-sulfur Analyzers JJG395	C:0.005%~0.010%	U=0.002%		
				C:0.010%~0.100%	U=0.005%		
				C:0.100%~1.000%	U=0.006%		
				C:1.00%~4.00%	U=0.03%		
				S:0.003%~0.010%	U=0.001%		
				S:0.010%~0.100%	U=0.002%		



No. CNAS L3031

第 6 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				S:0.100%~0.200%	$U=0.008\%$		
24	*Emission Spectrometer	Detection limit	V. R. of Emission Spectrometers JJG768	ICP:Zn $\leq 0.003\text{mg/L}$	$U_{\text{rel}}=13\%$		
				ICP:Ni $\leq 0.01\text{mg/L}$	$U_{\text{rel}}=14\%$		
				ICP:Mn $\leq 0.002\text{mg/L}$	$U_{\text{rel}}=14\%$		
				ICP:Cr $\leq 0.007\text{mg/L}$	$U_{\text{rel}}=13\%$		
				ICP:Cu $\leq 0.007\text{mg/L}$	$U_{\text{rel}}=11\%$		
				ICP:Ba $\leq 0.001\text{mg/L}$	$U_{\text{rel}}=13\%$		
				Direct reading spectrometer:C $\leq 0.005\%$	$U_{\text{rel}}=19\%$		
				Direct reading spectrometer:Si $\leq 0.005\%$	$U_{\text{rel}}=18\%$		
				Direct reading spectrometer:Mn $\leq 0.003\%$	$U_{\text{rel}}=18\%$		
				Direct reading spectrometer:Cr $\leq 0.003\%$	$U_{\text{rel}}=18\%$		
				Direct reading spectrometer:Ni $\leq 0.005\%$	$U_{\text{rel}}=19\%$		
				Direct reading spectrometer:V $\leq 0.001\%$	$U_{\text{rel}}=20\%$		
25	Automatic Potentiometric Titrators	mV	V. R. of Automatic Potentiometric Titrator JJG814	(-2000~2000)mV	$U=0.7\text{mV}$		
		Capacity		(1~10) mL	$U=0.003\text{mL}$		
26	*Frame Photometer	Detection limit	V. R. of Frame Photometer JJG630	K: $\leq 0.004\text{mmol/L}$	$U_{\text{rel}}=11\%$		
				Na: $\leq 0.008\text{mmol/L}$	$U_{\text{rel}}=12\%$		

No. CNAS L3031

第 7 页 共 14 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
27	*Fluorescence Spectrophotometer	Detection limit	V. R. of Fluorescence Spectrophotometers JJG537	$\leq 5 \times 10^{-10} \text{ g/mL}$	$U=1 \times 10^{-10} \text{ g/mL}$		
28	VOC Detector	Concentration	C. S. for Volatile Organic Compounds Photo Ionization Detectors JJF1172	$(1 \sim 1000) \times 10^{-6} \text{ mol/mol}$	$U_{\text{rel}}=3\%$		
29	*Sulfur Dioxide Gas Detector	Concentration	V. R. of Sulfur Dioxide Gas Detectors JJG551	$(1 \sim 500) \times 10^{-6} \text{ mol/mol}$	$U_{\text{rel}}=3\%$		
30	Ammonia Gas Detector	Concentration	V. R. of Ammonia Gas Detectors JJG1105	$(1 \sim 100) \times 10^{-6} \text{ mol/mol}$	$U_{\text{rel}}=3\%$		
31	Residual Chlorine Meter	Concentration	C. S. for Residual Chlorine Meter JJF1609	Total: $(0.02 \sim 5.00) \text{ mg/L}$	$U_{\text{rel}}=2\%$		
				Free: $(0.02 \sim 5.00) \text{ mg/L}$	$U_{\text{rel}}=3\%$		
32	Hand Saccharimeter(content-meter)and Hand Refractometer	Concentration	V. R. of Hand Saccharimeter(content-meter)and Hand Refractometer JJG820	$(1 \sim 80)\%$	$U_{\text{rel}}=2\%$		
		Refractivity		$1.3330 \sim 1.5200$	$U=8 \times 10^{-4}$		
33	*Atomic Fluorescence Spectrophotometer	Limited calculation	V. R. of Atomic Fluorescence Spectrophotometers JJG939	As: $\leq 0.4 \text{ ng}$	$U=0.04 \text{ ng}$		
				Sb: $\leq 0.4 \text{ ng}$	$U=0.03 \text{ ng}$		
34	Turbidimeters	Turbidity	V. R. of Turbidimeters JJG880	$(1 \sim 1000) \text{ NTU}$	$U_{\text{rel}}=4\%$		
35	Chemical Oxygen Demand(COD) Meters	Concentration	V. R. of Chemical Oxygen Demand(COD) Meters JJG975	$(1 \sim 1000) \text{ mg/L}$	$U_{\text{rel}}=3\%$		
36	*Total Organic Carbon Analyzer	Concentration	V. R. of Total Organic Carbon Analyzer JJG821	TOC: $(1 \sim 1000) \text{ mg/L}$	$U_{\text{rel}}=3\%$		
37	*Analyzers for Oil Content in Water	Concentration	V. R. of Analyzers for Oil Content in Water JJG 950	$(0 \sim 10) \text{ mg/L}$	$U=0.2 \text{ mg/L}$		
				$(10 \sim 1000) \text{ mg/L}$	$U_{\text{rel}}=3\%$		

No. CNAS L3031

第 8 页 共 14 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
38	*Melting-point Measurement Instruments	Temperature	V. R. of Melting-point Measurement Instruments JJG701	(50~300)°C	$U=0.2^{\circ}\text{C}$		
39	*Ozone Gas Analyzers	Concentration	V. R. of Ozone Gas Analyzers JJG 1077	(0.02~1.00) $\mu\text{mol/mol}$	$U_{\text{rel}}=3\%$		
				(1.00~400) $\mu\text{mol/mol}$	$U_{\text{rel}}=6\%$		
40	*Automatic Amoni Acid Analyzer	Detection limit	V. R. of Automatic Amoni Acid Analyzer JJG 1064	$\leq 1\text{nmol}$	$U_{\text{rel}}=14\%$		
41	*Micro Oxygen Analyzers	Concentration	Verification Regulation of Micro Oxygen Analyzers JJG945	$(1\sim 1000)\times 10^{-6}\text{mol/mol}$	$U_{\text{rel}}=3\%$		
42	*Determinators for Total Sulfur in Coal	Concentration	Verification Regulation of Determinators for Total Sulfur in Coal JJG1006	(0.01~1.00)%	$U=0.05\%$		
				(1.01~4.00)%	$U=0.07\%$		
				(4.01~6.00)%	$U=0.16\%$		
43	*Liquid chromatograph-Atomic Fluorescence Spectrometers	Minimum detecting limited concentration	Verification Regulation of Liquid chromatograph-Atomic Fluorescence Spectrometers JJG1151	As(V): $<1.0\text{ng}$	$U=0.02\text{ng}$		
				MMA: $<0.7\text{ng}$	$U=0.01\text{ng}$		
				DMA: $<0.7\text{ng}$	$U=0.01\text{ng}$		
44	*Flow Analyzers with Spectrophotography	Limited calculation	Calibration Specification for Flow Analyzers with Spectrophotography JJF1568	Cyanide: $\leq 0.002\text{ mg/L}$	$U=0.0003\text{mg/L}$		
				Volatile phenol in water: $\leq 0.002\text{ mg/L}$	$U=0.0002\text{mg/L}$		
				Hexavalent chromium: $\leq 0.004\text{ mg/L}$	$U=0.0004\text{mg/L}$		
				Sulfide: $\leq 0.005\text{ mg/L}$	$U=0.0006\text{mg/L}$		
				Total phosphorus: $\leq 0.01\text{ mg/L}$	$U=0.003\text{mg/L}$		

No. CNAS L3031

第 9 页 共 14 页



The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Total nitrogen: ≤ 0.04 mg/L	$U=0.004$ mg/L		
				Ammonia nitrogen: ≤ 0.04 mg/L	$U=0.005$ mg/L		
				Anion active detergent: ≤ 0.05 mg/L	$U=0.005$ mg/L		
3. Ionizing radiation							
1	*Medical Diagnostic X-ray Radiation Source	Air Kerma Rate	Verification Regulation of Medical Diagnostic X-ray Radiation Source JJG 744	0.1 mGy/min ~ 1 Gy/min	$U_{rel}=5\%$		
2	*Medical Diagnostic X-ray Radiation Source for Spiral Computed Tomography (CT)	CTDI	Verification Regulation of Medical Diagnostic X-ray Radiation Source for Spiral Computed Tomography (CT) JJG 961	0.1 mGy ~ 1 Gy	$U_{rel}=5\%$		
3	*Medical Diagnostic X-ray Radiation Source for Dental Panorama	Air Kerma Rate	Verification Regulation of Medical Diagnostic X-ray Radiation Source for Dental Panorama JJG 1101	0.1 mGy/min ~ 1 Gy/min	$U_{rel}=5\%$		
4	*X-ray Radiation Sources for Medical Computed Radiography System and Digital Radiography System	Air Kerma	Verification Regulation of X-ray Radiation Sources for Medical Computed Radiography System and Digital Radiography System JJG 1078	0.1 mGy ~ 100 mGy	$U_{rel}=4\%$		
5	*X-ray Security Inspection Equipment	Leakage of radiation dose	Calibration Specification for X-ray Security Inspection Equipment JJF 1275	1 μ Gy/h ~ 1 mGy/h	$U_{rel}=19\%$		

No. CNAS L3031

第 10 页 共 14 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
4.Optics							
1	Abbe Refractometer	refractive index	Verification Regulation of Abbe Refractometer JJG625	1.3000~1.7000	$U=7 \times 10^{-5}$		
5.Special calibration for Motor vehicles							
1	*Automobile Slip tester	Displacement	V.R.of Automobile Side Slip Tester JJG908	(-10~10) m/km	$U=0.1 \text{ m/km}$		
		Force		(50~150)N	$U=6\text{N}$		
2	*Axle (Wheel) Load Scales for Motor Vehicle Test	Mass	V.R.of Special Axle (Wheel) Load Scales for Motor Vehicle Test JJG1014	20kg~5t	$U_{\text{rel}}=0.3\%$		
3	*Roller Type Speedometer Tester	speed	V.R.of Roller Type Speedometer Tester JJG909	(20~100)km/h	$U_{\text{rel}}=1.4\%$		
		Length		(100~250)mm	$U_{\text{rel}}=0.26\%$		
4	*Vehicle Exhaust Emission Measuring Instruments	Gas Concentration	V. R. of Vehicle Exhaust Emission Measuring Instruments JJG688	HC: $(10 \sim 9999) \times 10^{-6}$	$U_{\text{rel}}=1.2\%$		
				CO: $(0.1 \sim 14.00) \times 10^{-2}$	$U_{\text{rel}}=1.8\%$		
				CO ₂ : $(0.1 \sim 18.0) \times 10^{-2}$	$U_{\text{rel}}=1.6\%$		
				NO: $(10 \sim 5000) \times 10^{-6}$	$U_{\text{rel}}=2.2\%$		
				O ₂ : $(0.1 \sim 25.0) \times 10^{-2}$	$U_{\text{rel}}=1.4\%$		
5	*Roller Opposite Forces Type Brake Tester	Force	V.R.of Roller Opposite Forces Type Brake Tester JJG906	(2000~30000)N	$U_{\text{rel}}=1.4\%$		
		adhesion coefficient		0.10~0.98	$U=0.08$		
6	*Headlamp Testers for Motor Vehicle	Luminous intensity	V.R.of Headlamp Tester for Motor Vehicle JJG745	(8000~60000) cd	$U_{\text{rel}}=6.3\%$		



No. CNAS L3031

第 11 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Offset of optical axis		up (0.1~20)cm/dam down(0.1~40)cm/dam left(0.1~40)cm/dam right(0.1~40)cm/dam	$U=1.2\text{cm/dam}$		
7	*Chassis Dynamometers for Automobile Emissions Testing	Speed	C.S for Chassis	(0~100)km/h	$U=0.1\text{ km/h}$		
		Torque	Dynamometers for Automobile Emissions	(5~2450)N	$U_{\text{rel}}=0.8\%$		
		Time	Testing JJF1221	(1~900) s	$U=1.4\text{s}$		
8	Motor Vehicle Tester fo Steering Force and Streering Angle	Torque	C.S of Motor Vehicle Testers for Steering Force and	(20~100) Nm	$U_{\text{rel}}=1.1\%$		
		Angel	Steering Angle JJF1196	Positive and negative direction:-1080°~1080°	$U=0.8^\circ$		
9	*Engine Speed Measuring Instruments	Rate	C.S for Motor Vehicle Engine Speed Measuring Instruments JJF1375	(500~6000) r/min	$U_{\text{rel}}=0.26\%$		
10	*retrotrflection coefficient meters for motor vehicle's reflecting Marking	retrotrflection coefficient	Calibration specification of retrotrflection coefficient meters for motor vehicle's reflecting Marking JJF 1747	white: (20~350) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$,	$U_{\text{rel}}=4.0\%$	合格评定 国家认可 委员会 认可证书专用章	
				yellow: (15~300) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$	$U_{\text{rel}}=4.0\%$		
				red: (10~60) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$	$U_{\text{rel}}=4.0\%$		
				green: (5~80) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$	$U_{\text{rel}}=4.0\%$		
				blue: (2~40) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$	$U_{\text{rel}}=4.0\%$		
				brown: (20~40) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$	$U_{\text{rel}}=4.0\%$		



No. CNAS L3031

第 12 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
11	*transmittance meter of automobile	transmittance	Calibration Specification for Transmittance Meter of Automobile JJF1225	(0~100)%	<i>U</i> =0.70%		
12	*the diesel vehicle nitrogen oxides measuring instruments	Gas Concentration	Calibration Specification for Diesel Vehicle Nitrogen Oxides (NO _x) Measuring Instruments JJF 1873	NO: (1~4000) ×10 ⁻⁶	<i>U</i> _{rel} =1.1%		
				NO ₂ : (1~1000) ×10 ⁻⁶	<i>U</i> _{rel} =3%		
				CO ₂ : (0.1~18.0) ×10 ⁻²	<i>U</i> _{rel} =1.7%		
13	*Flow Analyzer for Short Transient Loaded Mode of Gasoline Vehicles"	Gas flow	C.S of Flow Analyzer for Short Transient Loaded Mode of Gasoline Vehicles JJF1385	(95~180) L/s	<i>U</i> _{rel} =2.0%		
		Dilute mixture of Oxygen concentration		(5~21) %	<i>U</i> _{rel} =2.5%		
14	Manipulating Force Tester for Automotive Brake	Force value	Calibration Specification for Manipulating Force Tester for Automotive Brake JJF 1169	Pedal force meter: (100~1000) N	<i>U</i> _{rel} =0.6%		
				Handbreak force meter: (2~1000) N	<i>U</i> _{rel} =0.4%		
6.Special calibration for Medicine							
1	*Syringe Pump and Infusion Pump	Flow	Calibration Specification for Syringe Pumps and Infusion Pumps JJF1259	(5~20)mL/h	<i>U</i> _{rel} =2.6%		
				(20~200)mL/h	<i>U</i> _{rel} =1.4%		
				(200~1000)mL/h	<i>U</i> _{rel} =2.4%		
2	*Lung Ventilator	Tidal Volume	Calibration Specification for Lung Ventilators JJF1234	(0.2~1)L	<i>U</i> _{rel} =5%		
3	*Cardiac Defibrillator	Energy	Calibration Specification for Cardiac Defibrillators JJF1149	(2~50)J	<i>U</i> =2.7J		
				(50~360)J	<i>U</i> _{rel} =6%		
4	*ELISA Analytical Instrument	Absorbance	V. R. of ELISA Analytical Instrument JJG861	0.2~1.5	<i>U</i> =0.012		



No. CNAS L3031

第 13 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
5	*Pulmonary Function Measuring Instrument	vital capacity	Calibration Specification for the Pulmonary Function Measuring Instrument JJF 1213	(0.5~2)L	$U=0.024L$		
				(2~10)L	$U_{rel}=0.74\%$		
6	*Medical Centrifuges	rotational speed	Calibration Specification for Medical Centrifuges JJF 2004	(20~30000) r/min	$U_{rel}=0.04\%$		
7.Special calibration for Construction and transportation							
1	*Cement Mortar Mixer Complying	Rate	V.R.of The planet type mortar mixers JJF (JC) 123	(62~300)r/min	$U_{rel}=1.2\%$		
		Time		(20~130)s	$U_{rel}=1.3\%$		
		Length		(7~9)mm	$U_{rel}=0.6\%$		
2	*Mixer for Cement Paste	Rate	V.R.of Mixer for Cement Paste JJF (JC) 104	(62~300)r/min	$U_{rel}=1.2\%$		
		Time		(20~130)s	$U_{rel}=1.3\%$		
		Length		(5~6.5)mm	$U_{rel}=0.7\%$		



No. CNAS L3031



第 14 页 共 14 页

The scope of the accreditation in Chinese remains the definitive version.